

Content available at: <https://www.ipinnovative.com/open-access-journals>

IP International Journal of Medical Microbiology and Tropical Diseases

Journal homepage: <https://www.ijmmt.org/>

## Original Research Article

## Direct susceptibility testing by disk diffusion on positive BacT/ALERT blood cultures: A rapid and definite tool for antibiotic stewardship

V Lokeshwari<sup>1</sup>, Sandhya Bhat K<sup>1,\*</sup>, Vaishnava Devi<sup>2</sup><sup>1</sup>Dept. of Microbiology, Pondicherry Institute of Medical Sciences, Kalapet, Puducherry, India<sup>2</sup>Sri Lakshmi Narayana Institute of Medical Sciences, Pondicherry, India

## ARTICLE INFO

## Article history:

Received 30-04-2021

Accepted 29-07-2021

Available online 01-09-2021

## Keywords:

Blood stream infection

Blood culture

Direct susceptibility test

Disk diffusion

## ABSTRACT

**Introduction:** Blood stream infections continue to be the major cause of mortality and morbidity and hence early availability of direct susceptibility reports can be lifesaving. This study aims to ascertain if direct susceptibility testing (DST) can be used as a diagnostic tool in bacteremic patients and to correlate the results of both DST and standard antimicrobial susceptibility reports (AST), thereby serving to benefit both the patients and also to reduce the irrational use of antibiotics.

**Materials and Methods:** An experimental study was carried out after obtaining waiver of consent, in a tertiary care centre. A total of 37 patients were included in the study after careful consideration of the inclusion and exclusion criteria. Gram staining report, bacteriological profile, direct susceptibility report, antimicrobial susceptibility report of all the isolates were documented. Statistical analysis was done by using IBM SPSS software.

**Results:** Overall prevalence of sepsis was 40.5%. Gram negative bacteria were more commonly isolated (83.8) and *Escherichia coli* was the commonest isolate (51.4%). The antimicrobial resistance was observed maximum for amoxicillin/clavulanic acid (66.7%), ceftriaxone (60.6%), Cefotaxime (57.6%) and least for meropenem (9.1%), imipenem (6.1%). On comparison of DST with AST among 28 gram-negative Enterobacteriaceae isolates 15 minor errors (4.8%) and three major errors (0.97%) were recorded, with maximum errors being documented for piperacillin/tazobactam with five minor errors (17.9%) and one major error (3.6%).

**Conclusion:** DST is an important tool for early institution of targeted therapy and should be considered as one of the step towards antibiotic stewardship intervention.

This is an Open Access (OA) journal, and articles are distributed under the terms of the [Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License](https://creativecommons.org/licenses/by-nc-sa/4.0/), which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: [reprint@ipinnovative.com](mailto:reprint@ipinnovative.com)

## 1. Introduction

Blood stream infections (BSIs) represent an important cause of morbidity and mortality up to 20-50%.<sup>1</sup> The routine process of evaluation of patients with BSI involves, isolation of etiological agent by blood culture and antimicrobial susceptibility testing, which provides the susceptibility report for further prompt therapeutic interventions; however, these results will be available only with a delay of 48–72

hours after sampling.<sup>2,3</sup> Early administration of appropriate empirical antibiotic treatment has been associated with improved survival among patients with BSIs, yet up to 30-40% of patients with BSIs get inappropriate antibiotic treatment till the availability of susceptibility report.<sup>4-6</sup>

Decision on definitive therapy, whether there is need for escalation, or de-escalation, usually depends on susceptibility report from Microbiology laboratory.<sup>7</sup> Irrational antibiotic prescriptions may lead to immediate as well as long-term consequences such as the emergence of multidrug-resistant microorganisms and an increased risk of

\* Corresponding author.

E-mail address: [sandhyabhatk@gmail.com](mailto:sandhyabhatk@gmail.com) (S. Bhat K).

severe infections, morbidity and mortality; in turn leading to increased cost and prolonged hospitalization. Therefore, Microbiology laboratory should play a vital role to provide expedited reports on positive blood cultures, to guide the clinicians for antibiotic therapy, which may act as first step towards reduction of irrational use of broad spectrum antibiotics for therapy.<sup>3,5</sup>

Continuous monitoring by automated blood culture systems such as the BacT/ALERT lessens the duration of the time required for detection of positive blood cultures from bacteremic patients, often within 24 hours as compared to blind sub cultures from conventional blood culture systems. Presumptive identification of a positive blood culture is based on the Gram stain report of the positively flagged blood culture specimen. Following which, for about 30 to 45% of patients, the change of empirical treatment is being done.<sup>2,3,5,8</sup> Antimicrobial susceptibility testing (AST) routinely done by using standard method, takes about 24-48 hours to give the final results.<sup>9</sup> However, if direct susceptibility testing (DST) is being done by using the positive blood culture broth, the susceptibility report can be made available within 24 hours as compared to standard method, and this may have direct patient benefit in terms of initiation of prompt chemotherapy.<sup>10-12</sup> There may be extra benefits from the initiation of narrow the spectrum of antibiotic at an early stage. Which in turn may act as first step towards antibiotic stewardship. Hence this study was undertaken to determine the reliability of the results obtained by direct antimicrobial susceptibility testing in comparison to the results obtained by conventional standard antimicrobial susceptibility testing methods on positive blood culture samples.

## 2. Materials and Methods

An experimental study, based on diagnostic method of evaluation was conducted after obtaining waiver of consent from Institute Ethics Committee (RC.NO:RC/17/07) and necessary permission from the hospital management. All aerobic BacT/ALERT positive blood cultures from clinically suspected patients with blood stream infections (BSIs), between June 2017 to August 2017, from outpatient and inpatient departments of a tertiary care hospital were included in the study. Only positive blood culture specimens with a single organism in Gram stained smear report were included in the study. Repeat isolates from the same patients, positive blood specimens with more than one organism in Gram stained smear report or which yielded more than one isolate after subculture or which yielded fastidious bacteria and yeast were excluded from the study.

### 2.1. Brief procedure

The following procedures were followed on all aerobic BacT/ALERT positive blood culture specimens included in

the study:

1. Gram stained smear was made from the positive blood culture broth and the report was informed to the clinician telephonically.
2. Subculture from the BacT/ALERT positive bottle was done onto blood agar (BA), chocolate agar (CA) and MacConkey agar (MA) plates, and culture plates were incubated at 37<sup>0</sup>C.
3. If on Gram staining, gram-negative bacilli were detected, then direct susceptibility testing (DST) was performed using 2mL of positive blood culture broth<sup>3,8</sup> using antibiotic disks (HiMedia Laboratories, Mumbai, India) on Mueller Hinton agar (MHA). The antibiotic disks tested were: amoxicillin-clavulanate 20/10 $\mu$ g (Ac), gentamicin 10 $\mu$ g (G), amikacin 30 $\mu$ g (Ak), ciprofloxacin 5 $\mu$ g (Cf), levofloxacin 5 $\mu$ g (Le), ceftriaxone 30 $\mu$ g (Ci), cefotaxime 30 $\mu$ g (Ce), ceftazidime 30 $\mu$ g (CAZ), cefaperazone+sulbactam 75/10 $\mu$ g (Cfs), piperacillin-tazobactam 100/10 $\mu$ g (Pt), meropenem 10 $\mu$ g (Me), and imipenem 10 $\mu$ g (I). Direct bacterial identification was performed by inoculation of positive blood culture broth, using triple sugar iron agar (TSI), mannitol motility media, (MMM), citrate medium, and urea medium.<sup>13</sup>
4. If on Gram staining, gram-positive cocci (GPC) in clusters were detected, direct inoculation of positive blood culture broth (5 drops) [for direct tube coagulase test (DCT)], with 0.5 mL of human plasma at 35<sup>0</sup>C for 2 and 4 hours and clot formation was recorded.<sup>14</sup> Direct susceptibility testing for GPC clusters was performed using following antibiotic disks: penicillin 10U (P), gentamicin 10 $\mu$ g (G), amikacin 30 $\mu$ g (Ak), ciprofloxacin 5 $\mu$ g (Cf), ceftaxitin 30 $\mu$ g (Cn), linezolid 30 $\mu$ g (Lz). For screening methicillin and vancomycin resistance, oxacillin and vancomycin screen agar were used respectively, along with controls.<sup>9</sup>
5. Reading of sub cultured plates was done after 18hours of incubation and Gram stained smear was made from the colonies. Bacterial identification and antimicrobial susceptibility testing was done as per the standard guidelines.<sup>9,15</sup> The bacteriological profile of the organisms associated with bacteremia and their antibiotic susceptibility pattern was recorded (AST).
6. Direct susceptibility (DST) report was matched with conventional antimicrobial susceptibility report (AST) for reliability as per AST control document by U.S. Food and Drug Administration as follows-

### 2.2. Overall DST-AST match or mismatch

1. DST-AST Match was defined as complete match of DST report with that of AST report for all the antimicrobials tested.

2. DST-AST Mismatch: If DST report did not match with that of AST report for one or more antimicrobials tested.

### 2.3. Discrepancies between reports of individual antibiotics will be analyzed as per Food and Drug Administration (FDA) criteria (Table 1):<sup>16</sup>

1. (a) Minor discrepancy - The reference category result (conventional AST) is resistant (R) or sensitive (S) and the new device result (DST) is intermediate (I); or the reference result is I and the new device result is R or S.
- (b) Major discrepancy - The reference category result is S and the new device result is R
- (c) Very major discrepancy- The reference category result is R and the new device result is S.

### 2.4. Methods of statistical analysis


Data entry was done using Microsoft Excel and analysis was done using SPSS for Windows Version SPSS version 21.0 (SPSS Inc., Chicago, IL, USA). Mean and standard deviations were calculated for numerical variables. Percentages were calculated for categorical variables. Correlation between DST to AST was analyzed using Pearson correlation and two tailed P value of <0.001 was considered significant.

## 3. Results

Out of 773 clinically suspected patients with bacteremia admitted in various wards of a tertiary care center, 74 (9.6%) were blood culture positive, during the study period. Out of these patients 37 (50%) were included for the study and the remaining were excluded from the study, as they had grown fastidious organisms, yeast and skin contaminants from positive blood cultures. Maximum patients were in the age group of 40-60 years (40.5%) and the mean age of the study population was  $46.7 \pm 18.8$  years. Out of total 37 patients, majority [15 (40.5%)] were clinically diagnosed to have sepsis and septic shock.


Gram staining report of the smear prepared from the positive BacT/ALERT bottles revealed 33 gram-negative rods (89.2%) and four gram-positive cocci in clusters (10.8%). Among the gram-negative bacteria, *Escherichia coli* was the commonest isolate (51.4%), followed by *Klebsiella pneumoniae* (10.8%), *Aeromonas caviae* (8.1), *Pseudomonas aeruginosa*, *Salmonella Typhi* and *Salmonella Paratyphi A* (5.4% each) and *Enterobacter spp.* (2.7%). Among the gram-positive bacteria, *Staphylococcus aureus* was isolated from four cases, of which two were methicillin resistant *Staphylococcus aureus* (50%).

Overall antimicrobial resistance rate among gram-negative bacterial isolates (n=33) detected by direct susceptibility testing has been shown in the Figure 1.



**Fig. 1:** Antimicrobial agents tested for gram-negative isolates by DST method and their resistance pattern (n=33).

Among 33 gram-negative bacterial isolates, 28 (75.7%) were Enterobacteriaceae, out of which four were *Salmonella* species. The overall antimicrobial resistance rate among Enterobacteriaceae (except for *Salmonella*) by AST method has been shown in Figure 2. All four *Salmonella* isolates were uniformly susceptible to ceftriaxone, and two of *Salmonella Paratyphi A* isolates were resistant to ciprofloxacin and two of *Salmonella Typhi* isolates were intermediate susceptible to ciprofloxacin. However other tested antimicrobial agents same as tested for other Enterobacteriaceae were not reported as they are not recommended for *Salmonella* isolates by CLSI guidelines.<sup>9</sup> Cefazidime susceptibility report was not analyzed for Enterobacteriaceae isolates.



**Fig. 2:** Antimicrobial agents tested (n= 11) for Enterobacteriaceae isolates (except *Salmonella*) by AST method and their resistance pattern (n=24)

Among the non-Enterobacteriaceae isolates (n=5), overall antibiotic resistance in decreasing order was as follows: cefotaxime (83.3%), ciprofloxacin and ceftazidime (40% each), cefoperazone/sulbactam, amikacin, levofloxacin, meropenem and imipenem (20% each),

**Table 1:** Food and Drug Administration (FDA) criteria for analyzing discrepancies between reports of AST and DST

| Reference method (AST) | Minor discrepancy |        | Major discrepancy | Very major discrepancy |
|------------------------|-------------------|--------|-------------------|------------------------|
|                        | R or S            | I      | S                 | R                      |
| New method (DST)       | I                 | R or S | R                 | S                      |

and no resistance was observed for piperacillin/tazobactam and gentamicin. Amoxicillin–clavulanate and ceftriaxone 30 $\mu$ g susceptibility reports were not analyzed for non Enterobacteriaceae isolates.

Among the four gram-positive bacterial isolates, overall antimicrobial resistance rate by both DST and AST methods, in the decreasing order was as follows: penicillin (100%), cefoxitin (50%), ciprofloxacin, erythromycin, cotrimoxazole, gentamicin, amikacin and clindamycin (25% each) and no resistance was observed to vancomycin, and linezolid.

Comparison of interpretative results with direct susceptibility testing method (DST) and standard antimicrobial susceptibility testing method (AST) for Enterobacteriaceae is shown in Table 2. In this study we found total of 15 minor errors (4.8%) and three major errors (0.97%) for 28 Enterobacteriaceae isolates against 11 antimicrobial agents tested (i.e., total of 308 isolate/antibiotic combinations). Maximum errors were recorded for piperacillin/tazobactam with five (17.9%) minor errors and one (3.6%) major error. However, we did not find any errors for gentamicin, ceftriaxone, cefotaxime, levofloxacin, and meropenem.

Among non Enterobacteriaceae isolates (n=5), comparison of interpretative results with DST and AST, we found total of two minor errors, one each for gentamicin and imipenem; however, we did not find any very major error. (i.e., when organisms were found to be susceptible by DST and actually resistant by standard AST method).

In the present study, overall minor errors observed among all 33 gram-negative rods was 17 (4.7%, i.e., 17/363 X100), out of total of 358 organism-antimicrobial agent combinations, [i.e (28 Enterobacteriaceae isolates X 11 antibiotics tested) + (5 non Enterobacteriaceae isolates X 10 antibiotics tested)].<sup>16</sup> Among four of Staphylococcus aureus isolates we did not find any mismatch in the susceptibility interpretation by DST and AST methods.

Overall agreement between susceptibility report by DST and AST ranged from 75% to 100%. Maximum agreement of 100% was observed for gentamicin, ceftriaxone, meropenem, levofloxacin, and least agreement (75%) was observed for piperacillin/tazobactam with kappa value of agreement of 0.427 and p value of 0.001, moderate level of agreement.<sup>17</sup> The level of agreement (kappa value with p value) between DST result with standard AST method for all tested antimicrobial agents for Enterobacteriaceae isolates has been shown in Table 3.

#### 4. Discussion

Despite of advances in diagnosis and treatment in the medical care, bacterial sepsis remains as one of the leading cause of morbidity and mortality, particularly among neonates and elderly patients in developing countries. The etiological agents causing sepsis and their antimicrobial susceptibility are constantly evolving. Hence the study of bacteriological profile with direct antibiotic susceptibility testing may play an important role in effective early management of blood stream infection cases.<sup>4,12,18</sup>

In the current study, overall prevalence of bacteremia was 9.6% and majority of patients (40.5%) were clinically diagnosed to have sepsis and septic shock. In the current study, by direct gram staining of positive blood culture BacT/ALERT bottles gram negative bacteria were more common (89.1%) than gram-positive bacteria (10.9%). Notification of direct Gram stain reports of positive blood cultures have a significantly higher impact than the release of antimicrobial susceptibility report for empirical antimicrobial management of BSIs.<sup>19,20</sup>

Of the 33 gram-negative bacterial isolates, Enterobacteriaceae were majority (n= 28, 75.7%) and among which, Escherichia coli was the commonest isolate (51.4%), followed by Klebsiella pneumoniae (10.8%). Among the gram-positive bacteria, four Staphylococcus aureus were isolated, of which two were (50%) methicillin resistant Staphylococcus aureus (MRSA). In various other studies by Rajshekar D et al., and Annamallaei et al., similar findings were documented.<sup>19,21</sup> However, as our sample size was smaller (37 only), we cannot further opine on the common bacteriological profile.

Antimicrobial resistance rate among Enterobacteriaceae isolates by standard AST method was observed highest with cefotaxime (83.3%), ceftriaxone and amoxicillin/clavulanic acid (79.2% each) followed by levofloxacin (58.3%), ciprofloxacin and gentamicin (45.8% each), amikacin (29.2%), cefoperazone/sulbactam, piperacillin/tazobactam (25% each), least resistance was observed for meropenam (8.3%), and imipenem (4.2%) respectively. Similar observations were made by various other studies by Negussie A et al, Annamallaei et al., and Rajeevan et al.<sup>18,21,22</sup> Among non Enterobacteriaceae isolates resistance was observed highest for cefotaxime (83.3%), and least resistance was observed for meropenem and imipenem (20%). As these isolates were few (only five), hence any interpretation of resistance would not be correct; which requires larger sample size. Among the four Staphylococcus aureus isolates, two were MRSA. All four isolates were

**Table 2:** Comparison of interpretative results with DST and AST for Enterobacteriaceae isolates

| Antimicrobial agents tested | Minor error (%) | Major error (%) | Very major error (%) |
|-----------------------------|-----------------|-----------------|----------------------|
| Ciprofloxacin               | 3 (10.7)        | 0               | 0                    |
| Amoxicillin/clavulanic acid | 2 (7.1)         | 0               | 0                    |
| Gentamicin                  | 0               | 0               | 0                    |
| Amikacin                    | 2 (7.1)         | 1 (3.6)         | 0                    |
| Ceftriaxone                 | 0               | 0               | 0                    |
| Cefotaxime                  | 0               | 1 (3.6)         | 0                    |
| Cefaperazone/sulbactam      | 2 (7.1)         | 0               | 0                    |
| Piperacillin/tazobactam     | 5 (17.9)        | 1 (3.6)         | 0                    |
| Imipenem                    | 1 (3.6)         | 0               | 0                    |
| Levofloxacin                | 0               | 0               | 0                    |
| Meropenem                   | 0               | 0               | 0                    |
| Total                       | 15 (4.8)        | 3 (0.97)        | 0                    |

**Table 3:** Agreement between DST result with standard AST method for all tested antimicrobial agents for Enterobacteriaceae isolates (n=24, except Salmonella isolates)

| Antimicrobial agents tested | Total disagreement (n=24) | Total agreement (n=24) | Agreement in % | Kappa value of agreement | P value |
|-----------------------------|---------------------------|------------------------|----------------|--------------------------|---------|
| Ciprofloxacin               | 3                         | 21                     | 88             | 0.777                    | 0.000** |
| Amoxicillin/Clavulanic acid | 2                         | 22                     | 92             | 0.847                    | 0.000** |
| Gentamicin                  | 0                         | 24                     | 100            | 0.930                    | 0.000** |
| Amikacin                    | 3                         | 21                     | 88             | 0.777                    | 0.000** |
| Ceftriaxone                 | 0                         | 24                     | 100            | 1.000                    | 0.000** |
| Cefotaxime                  | 1                         | 23                     | 96             | 0.901                    | 0.000** |
| Cefaperazone/sulbactam      | 2                         | 22                     | 92             | 0.847                    | 0.000** |
| Piperacillin/tazobactam     | 6                         | 18                     | 75             | 0.427                    | 0.001*  |
| Imipenem                    | 1                         | 23                     | 96             | 0.901                    | 0.000** |
| Levofloxacin                | 0                         | 24                     | 100            | 0.935                    | 0.000** |
| Meropenem                   | 0                         | 24                     | 100            | 1.000                    | 0.000** |

\*P value of 0.000 is highly significant,

\*\*P value of 0.001 is significant.

penicillin resistant and uniformly susceptible to vancomycin and linezolid. Studies done by many other researchers also observed same susceptibility pattern for gram-positive bacterial isolates.<sup>1,19,23</sup>

Direct susceptibility report may reduce the turnaround time and facilitate the early initiation of appropriate antimicrobial agent as compared to conventional susceptibility testing method; however, there should be higher level of agreement between the susceptibility reports of DST when compared to the standard AST reports.<sup>3,4,24</sup> In this study we found total of 15 minor errors (4.8%) and three major errors (0.97%) for 28 Enterobacteriaceae isolates, out of total of 308 antimicrobial agent-organism combinations tested. This means, zone diameters by standard method were greater than DST and maximum errors were found for piperacillin/tazobactam with five (17.9%) minor errors and one (3.6%) major error. However, we did not find any errors for gentamicin, ceftriaxone, cefotaxime, levofloxacin, and meropenem. As in this study non-Enterobacteriaceae isolates were few (n=5), we found

two minor errors, one each for gentamicin and imipenem and we did not find any major errors. Similar observations were made by study by Noman F et al., Coorevits L et al and Rajshekar D et al. Most of the discrepancies were in beta lactam/beta lactamase combinations and aminoglycosides.<sup>3,12,19</sup>

While reporting DST results of these antibiotics one should be cautious, as these are major antibiotics used for life threatening infections. The main sources of errors by DST may be i) mixed cultures or ii) non-standardized inoculum size, iii) disk to disk variations. As in the present study, specimens yielding more than one bacteria were excluded from analysis, and our data are based on mono-microbial positive blood cultures, first source of error may be excluded. We did not observe any discrepancies in interpretation of susceptibility report for gram-positive bacteria, considering interpretation of methicillin resistance in staphylococci, by cefoxitin disk and oxacillin screen agar testing. Similar findings were observed in other studies.<sup>19,25</sup>

Currently with availability of reference documents (CLSA M 47-A and EUCAST RAST method v 1.1), recommending the standard procedures for performing direct susceptibility from the positive blood culture broth and also keeping in mind the advantage of early reporting in management of critically ill patients with sepsis, we should consider DST as effective step towards antimicrobial stewardship to control the irrational use of broad spectrum antibiotics. This may further contribute to reduce the antibiotic resistance; however large size studies need to be conducted to document these observations.

## 5. Conclusion

Early recognition of sepsis and appropriate microbiological and radiological investigations to aid in diagnosis is essential to improve the patient outcome. When used selectively and interpreted cautiously DST reports from positive blood culture samples is potentially useful in management of critically ill patients. However, DST report must always be confirmed by the standard AST report. In the past, due to unavailability of standard recommended method DST results have reported erroneous results, but however recently both CLSI and EUCAST guidelines have come up with the standard procedure for performing DST on positive blood culture sample. However, larger sample size would have helped us to get conclusive evidence on impact of definitive therapy on outcome of patients with sepsis.

There is a high concordance between DST and conventional AST. Considering the clinical and economic benefits of early availability of reports by DST, a close coordination between the laboratory and clinicians for interpretation of microbiological reports are crucial for management of patients with BSIs.

## 6. Source of Funding

This project was supported and funded by PIMS fellowship 2017.

## 7. Conflicts of interest

There are no conflicts of interest.

## References

- Gupta S, Kashyap B. Bacteriological profile and antibiogram of blood culture isolates from a tertiary care hospital of North India. *Trop J Med Res.* 2016;19(2):94.
- Kirn TJ, Weinstein MP. Update on blood cultures: how to obtain, process, report, and interpret. *Clin Microbiol Infect.* 2013;19(6):513–20.
- Coorevits L, Boelens J, Claeys G. Direct susceptibility testing by disk diffusion on clinical samples: a rapid and accurate tool for antibiotic stewardship. *Eur J Clin Microbiol Infect Dis.* 2015;34(6):1207–12.
- Lahanas S, Lee A, Menon V, Janto C. Utility of direct susceptibility testing on blood cultures: is it still worthwhile. *J Med Microbiol.* 2016;65(6):501–9.
- Sogaard M, Norgaard M, Schonheyder HC. First Notification of Positive Blood Cultures and the High Accuracy of the Gram Stain Report. *J Clin Microbiol.* 2007;45(4):1113–7.
- Bouza E, Sousa D, Muñoz P, Rodríguez-Crèixems M, Fron C, Lechuz JG. Bloodstream infections: a trial of the impact of different methods of reporting positive blood culture results. *Clin Infect Dis Off Publ Infect Dis Soc Am.* 2004;39(8):1161–9.
- Sahm DF, Boonlayangoor S, Morello JA. Direct susceptibility testing of blood culture isolates with the AutoMicrobic System (AMS). *Diagn Microbiol Infect Dis.* 1987;8(1):1–11.
- Waites KB, Brookings ES, Moser SA, Zimmer BL. Direct susceptibility testing with positive Bact/Alert blood cultures by using MicroScan overnight and rapid panels. *J Clin Microbiol.* 1998;36(7):2052–6.
- Clinical Laboratory Standards Institute. Performance Standards for Antimicrobial Susceptibility Testing. 27th ed. CLSI Document M100-S22. Wayne, PA: CLSI; 2017.
- Andrews JM. for the BSAC Working Party on Susceptibility Testing. BSAC standardized disc susceptibility testing method (version 6). *J Antimicrob Chemother.* 2007;60(1):20–41.
- The European Committee on Antimicrobial Susceptibility Testing. Breakpoint tables for interpretation of MICs and zone diameters, version 10.0, 2018" (or another relevant version and year);.
- Noman F, Jehan A, Ahmed A. Reliability of direct sensitivity determination of blood cultures. *J Coll Physicians Surg Pak.* 2008;18(10):660–1.
- Yu FL, Lin MH, Lee JC, Lian YL, Lin CW, Chen CT, et al. Comparison of Antimicrobial Susceptibility Testing of Isolates from Blood Cultures by Direct Inoculation Method and PHOENIX. *J Biomed Lab Sci.* 2011;23(1):23.
- Qian Q, Eichelberger K, Kirby JE. Rapid Identification of Staphylococcus aureus in Blood Cultures by Use of the Direct Tube Coagulase Test. *J Clin Microbiol.* 2007;45(7):2267–9.
- Mackie TJ, McCartney JE. McCartney JE. Practical Medical Microbiology. In: 14th Edn.. vol. 21. New York: Churchill Livingstone; 1996. p. 385–402.
- Health C for D and R. Search for FDA Guidance Documents - Class II Special Controls Guidance Document: Antimicrobial Susceptibility Test (AST) Systems [Internet]. [cited 2017 Dec 13]. Available from: <https://www.fda.gov/RegulatoryInformation/Guidances/ucm080564.htm>.
- Cohen's kappa. In: Wikipedia [Internet]; 2017. Available from: [https://en.wikipedia.org/w/index.php?title=Cohen%27s\\_kappa&oldid=815294345](https://en.wikipedia.org/w/index.php?title=Cohen%27s_kappa&oldid=815294345).
- Negussie A, Mulugeta G, Bedru A, Ali I, Shimeles D, Lema T, et al. Bacteriological Profile and Antimicrobial Susceptibility Pattern of Blood Culture Isolates among Septicemia Suspected Children in Selected Hospitals Addis Ababa, Ethiopia. *Int J Biol Med Res.* 2015;6(1):4709–17.
- Rajshekar D, Chaudhari KV, Bhat P, Prakash SS, Raghvan R, Vasanth S, et al. Evaluation of performance of direct disk diffusion test from positively flagged blood culture broth: A large scale study from South India. *J Lab Physicians.* 2019;11(2):154–60. doi:10.4103/JLP.JLP\_137\_18.
- Mathur P, Gupta B, Behera B. Blood culture gram stain, acridine orange stain and direct sensitivity-based antimicrobial therapy of bloodstream infection in patients with trauma. *Indian J Med Microbiol.* 2010;28(2):138–42. doi:10.4103/0255-0857.62491.
- Annamallaei S, Bhat KS. Correlation of empiric antibiotic use with susceptibility pattern of blood isolates in septicemic patients in an Intensive Care Unit. *J Curr Res Sci Med.* 2017;3(1):29–35.
- Rajeevan S, Ahmad SM, Jasmin PT. Study of prevalence and antimicrobial susceptibility pattern in blood isolates from a tertiary care hospital in North Kerala, India. *Int J Curr Microbiol App Sci.* 2014;3(4):655–62.
- Beuving J, Wolffs P, Hansen WLJ, Stobberingh EE, Bruggeman CA, Kessels A. Impact of same-day antibiotic susceptibility testing on time to appropriate antibiotic treatment of patients with bacteraemia: a randomised controlled trial. *Eur J Clin Microbiol Infect Dis.* 2015;34(4):831–8.

24. Cueto MD, Ceballos E, Martinez-Martinez L, Perea EJ, Pascual A. Use of Positive Blood Cultures for Direct Identification and Susceptibility Testing with the Vitek 2 System. *J Clin Microbiol.* 2004;42(8):3734–8.
25. Bennett K, Sharp SE. Rapid differentiation of methicillin-resistant *Staphylococcus aureus* and methicillin-susceptible *Staphylococcus aureus* from blood cultures by use of a direct ceftioxin disk diffusion test. *J Clin Microbiol.* 2008;46(11):3836–8.

**Sandhya Bhat K**, Professor

**Vaishnava Devi**, Tutor

### **Author biography**

**V Lokeshwari**, Intern

**Cite this article:** Lokeshwari V, Bhat K S, Devi V. Direct susceptibility testing by disk diffusion on positive BacT/ALERT blood cultures: A rapid and definite tool for antibiotic stewardship. *IP Int J Med Microbiol Trop Dis* 2021;7(3):179-185.